

Equipment Technical Bulletin

P-6100T Pump Stalling/PCB Component Failure



PROBLEM

Customers are reporting that P-6100T pump motors are stalling and components on the PCB are failing when the unit is used with tubing other than Silicone

EXPLANATION

P-6100T is designed for use with Silicone tubing exclusively. **Use of any other tubing material with P-6100T voids the warranty.**

SOLUTION

Always use Silicone tubing with P-6100T. If PCB fails as a result of using a different tubing material, you will need to replace the PCB.

For applications requiring tubing material other than Silicone, please consider using our PR1100 or LS3000 Series pumps.



Beta Technology
2841 Mission Street
Santa Cruz • CA
U.S.A. • 95060-2142

TEL • 831 • 426 • 0882
TEL • 800 • 858 • 2382
FAX • 838 • 423 • 4573
FAX • 800 • 221 • 8416

<http://www.beta-technology.com>
sales@beta-technology.com

Beta Technology is ISO9001 Certified